

a final draft of the guidance was submitted to the ICH Steering Committee and endorsed by the three participating regulatory agencies in July 2000.

The guidance sets forth critical issues in pediatric drug development and approaches to the safe, efficient, and ethical study of medicinal products in the pediatric population. The guidance addresses the following clinical study issues: (1) Considerations when initiating a pediatric program for a medicinal product; (2) timing of initiation of pediatric studies during medicinal product development; (3) types of studies (pharmacokinetic, pharmacokinetic/pharmacodynamic, efficacy, safety); (4) age categories for studies; and (5) ethics of pediatric clinical investigation. The guidance is not comprehensive, but is intended to be used in conjunction with other ICH guidances and documents from regional regulatory authorities and pediatric societies. The guidance is intended to encourage and facilitate the timely development of pediatric medicinal products internationally.

This guidance represents the agency's current thinking on clinical investigation of medicinal products in the pediatric population. It does not create or confer any rights for or on any person and does not operate to bind FDA or the public. An alternative approach may be used if such approach satisfies the requirements of the applicable statutes, regulations, or both.

Interested persons may submit to the Dockets Management Branch (address above) written comments on the guidance at any time. Two copies of any comments are to be submitted, except that individuals may submit one copy. Comments are to be identified with the docket number found in brackets in the heading of this document. The guidance and received comments may be seen in the Dockets Management Branch between 9 a.m. and 4 p.m., Monday through Friday.

Electronic Access

Copies of the guidance are available on the Internet at <http://www.fda.gov/cder/guidance/index.htm> or <http://www.fda.gov/cber/publications.htm>.

Dated: December 7, 2000.

Margaret M. Dotzel,

Associate Commissioner for Policy.

[FR Doc. 00-31947 Filed 12-14-00; 8:45 am]

BILLING CODE 4160-01-F

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. 00P-1554]

Medical Devices; Exemptions From Premarket Notification; Class II Devices

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA) is publishing a notice announcing that it has received a petition requesting exemption from the premarket notification requirements for pharmacy compounding systems classified within the intravascular administration set, a class II device (special controls). FDA is publishing this notice in order to obtain comments on this petition in accordance with procedures established by the Food and Drug Administration Modernization Act of 1997 (FDAMA).

DATES: Submit written comments by January 16, 2001.

ADDRESSES: Submit written comments on this notice to the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852.

FOR FURTHER INFORMATION CONTACT: Heather S. Rosecrans, Center for Devices and Radiological Health (HFZ-404), Food and Drug Administration, 9200 Corporate Blvd., Rockville, MD 20850, 301-594-1190.

SUPPLEMENTARY INFORMATION:

I. Statutory Background

Under section 513 of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360c), FDA must classify devices into one of three regulatory classes: Class I, class II, or class III. FDA classification of a device is determined by the amount of regulation necessary to provide a reasonable assurance of safety and effectiveness. Under the Medical Device Amendments of 1976 (the 1976 amendments (Public Law 94-295)), as amended by the Safe Medical Devices Act of 1990 (the SMDA) (Public Law 101-629)), devices are to be classified into class I (general controls) if there is information showing that the general controls of the act are sufficient to ensure safety and effectiveness; into class II (special controls), if general controls, by themselves, are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls to provide such

assurance; and into class III (premarket approval), if there is insufficient information to support classifying a device into class I or class II and the device is a life-sustaining or life-supporting device or is for a use that is of substantial importance in preventing impairment of human health, or presents a potential unreasonable risk of illness or injury.

Most generic types of devices that were on the market before the date of the 1976 amendments (May 28, 1976) (generally referred to as preamendments devices) have been classified by FDA under the procedures set forth in section 513(c) and (d) of the act through the issuance of classification regulations into one of these three regulatory classes. Devices introduced into interstate commerce for the first time on or after May 28, 1976 (generally referred to as postamendments devices), are classified through the premarket notification process under section 510(k) of the act (21 U.S.C. 360(k)). Section 510(k) of the act and the implementing regulations (21 CFR part 807) require persons who intend to market a new device to submit a premarket notification report containing information that allows FDA to determine whether the new device is "substantially equivalent" within the meaning of section 513(i) of the act to a legally marketed device that does not require premarket approval.

On November 21, 1997, the President signed into law FDAMA (Public Law 105-115). Section 206 of FDAMA, in part, added a new section 510(m) to the act. Section 510(m)(1) of the act requires FDA, within 60 days after enactment of the FDAMA, to publish in the **Federal Register** a list of each type of class II device that does not require a report under section 510(k) of the act to provide reasonable assurance of safety and effectiveness. Section 510(m) of the act further provides that a 510(k) will no longer be required for these devices upon the date of publication of the list in the **Federal Register**. FDA published that list in the **Federal Register** of January 21, 1998 (63 FR 3142). In the **Federal Register** of November 3, 1998 (63 FR 59222), FDA published a final rule codifying these exemptions.

Section 510(m)(2) of the act provides that, 1 day after date of publication of the list under section 510(m)(1), FDA may exempt a device on its own initiative or upon petition of an interested person, if FDA determines that a 510(k) is not necessary to provide reasonable assurance of the safety and effectiveness of the device. This section requires FDA to publish in the **Federal Register** a notice of intent to exempt a

device, or of the petition, and to provide a 30-day comment period. Within 120 days of publication of this document, FDA must publish in the **Federal Register** its final determination regarding the exemption of the device that was the subject of the notice. If FDA fails to respond to a petition under this section within 180 days of receiving it, the petition shall be deemed granted.

II. Criteria for Exemption

There are a number of factors FDA may consider to determine whether a 510(k) is necessary to provide reasonable assurance of the safety and effectiveness of a class II device. These factors are discussed in the guidance the agency issued on February 19, 1998, entitled "Procedures for Class II Device Exemptions From Premarket Notification, Guidance for Industry and CDRH Staff." That guidance can be obtained through the Internet on the CDRH home page at <http://www.fda.gov/cdrh> or by facsimile through CDRH Facts-on-Demand at 1-800-899-0381 or 301-827-0111. Specify "159" when prompted for the document shelf number.

III. Petition

FDA received the following petition requesting an exemption from premarket notification for class II devices: Baxter Healthcare, Pharmacy Compounding Systems Classified within the *Intravascular administration set*, 21 CFR 880.5440.

IV. Comments

Interested persons may submit to the Dockets Management Branch (address above) written comments regarding this petition by January 16, 2001. Two copies of any comments are to be submitted, except that individuals may submit one copy. Comments are to be identified with the docket number found in brackets in the heading of this document. The petition and received comments may be seen in the Dockets Management Branch between 9 a.m. and 4 p.m., Monday through Friday.

Dated: December 5, 2000.

Linda S. Kahan,

Deputy Director for Regulations Policy, Center for Devices and Radiological Health.

[FR Doc. 00-31961 Filed 12-14-00; 8:45 am]

BILLING CODE 4160-01-F

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket Nos. 00M-1391, 00M-1536, 00M-1447, 00M-1522, 00M-0809, 00M-1517, 00M-1451, 00M-1448, 00M-1507, 00M-1389, 00M-1388, 00M-1508, 00M-1390, 00M-1386, 00M-1387, 00M-1414, 00M-1415, 00M-1416, 00M-1495, 00M-1437, 00M-1475, 00M-1483, 00M-1515, 00M-1524, 00M-1523]

Medical Devices; Availability of Safety and Effectiveness Summaries for Premarket Approval Applications

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA) is publishing a list of premarket approval applications (PMA's) that have been approved. This list is intended to inform the public of the availability of safety and effectiveness summaries of approved PMA's through the Internet and the agency's Dockets Management Branch.

ADDRESSES: Submit a written request for copies of summaries of safety and effectiveness to the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Please cite the appropriate docket number as listed in table 1 of this document when submitting a written request. See the **SUPPLEMENTARY INFORMATION** section for electronic access to the summary of safety and effectiveness

FOR FURTHER INFORMATION CONTACT: Tinh Nguyen, Center for Devices and Radiological Health (HFZ-402), Food and Drug Administration, 9200 Corporate Blvd., Rockville, MD 20850, 301-594-2186.

SUPPLEMENTARY INFORMATION:

I. Background

In the **Federal Register** of January 30, 1998 (63 FR 4571), FDA published a final rule to revise §§ 814.44(d) and 814.45(d) (21 CFR 814.44(d) and 814.45(d)) to discontinue publication of individual PMA approvals and denials in the **Federal Register**. Instead, revised §§ 814.44(d) and 814.45(d) state that FDA will notify the public of PMA

approvals and denials by posting them on FDA's home page at <http://www.fda.gov> on the Internet; by placing the summaries of safety and effectiveness on the Internet and in FDA's Dockets Management Branch; and by publishing in the **Federal Register** after each quarter a list of available safety and effectiveness summaries of approved PMA's and denials announced in that quarter.

FDA believes that this procedure expedites public notification of these actions because announcements can be placed on the Internet more quickly than they can be published in the **Federal Register**, and FDA believes that the Internet is accessible to more people than the **Federal Register**.

In accordance with section 515(d)(4) and (e)(2) of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360e(d)(4) and (e)(2)), notification of an order approving, denying, or withdrawing approval of a PMA will continue to include a notice of opportunity to request review of the order under section 515(g) of the act. The 30-day period for requesting reconsideration of an FDA action under § 10.33(b) (21 CFR 10.33(b)) for notices announcing approval of a PMA begins on the day the notice is placed on the Internet. Section 10.33(b) provides that FDA may, for good cause, extend this 30-day period. Reconsideration of a denial or withdrawal of approval of a PMA may be sought only by the applicant; in these cases, the 30-day period will begin when the applicant is notified by FDA in writing of its decision.

The following is a list of approved PMA's for which summaries of safety and effectiveness were placed on the Internet in accordance with the procedure explained previously from July 1, 2000, through September 30, 2000. There were no denial actions during this period. The list provides the manufacturer's name, the product's generic name or the trade name, and the approval date.