

DEPARTMENT OF LABOR**Employment and Training
Administration****[TA-W-53,364]****Teleflex Automotive Group, Warren,
Michigan; Notice of Termination of
Investigation**

Pursuant to Section 221 of the Trade Act of 1974, as amended, an investigation was initiated on October 28, 2003 in response to a petition filed on behalf of workers at Teleflex Automotive Group, Warren, Michigan. The workers produced automotive cable assembly components.

All workers were separated from the subject firm more than one year before the date of the petition. Section 223 (b) of the Act specifies that no certification may apply to any worker whose last separation occurred more than one year before the date of the petition. Consequently, further investigation in this case would serve no purpose, and the investigation has been terminated.

Signed at Washington, DC, this 13th day of November 2003.

Linda G. Poole,

*Certifying Officer, Division of Trade
Adjustment Assistance.*

[FR Doc. 03-30089 Filed 12-2-03; 8:45 am]

BILLING CODE 4510-30-P

DEPARTMENT OF LABOR**Employment and Training
Administration****[TA-W-53,263]****Thomson, Inc., Indianapolis, IN; Notice
of Termination of Investigation**

Pursuant to section 221 of the Trade Act of 1974, as amended, an investigation was initiated on October 16, 2003 in response to a petition filed by International Brotherhood of Electrical Workers, Local 1048, on behalf of workers at Thomson, Inc., Indianapolis, Indiana.

The petitioner has requested that the petition be withdrawn. Consequently, the investigation has been terminated.

Signed at Washington, DC, this 24th day of October 2003.

Richard Church,

*Certifying Officer, Division of Trade
Adjustment Assistance.*

[FR Doc. 03-30085 Filed 12-2-03; 8:45 am]

BILLING CODE 4510-30-P

DEPARTMENT OF LABOR**Employment and Training
Administration****[TA-W-53,242]****Wellington Leisure Products, Leesville
Synthetic Fibers, Leesville, SC; Notice
of Termination of Investigation**

Pursuant to section 221 of the Trade Act of 1974, as amended, an investigation was initiated on October 14, 2003 in response to a petition filed by a company official on behalf of workers at Wellington Leisure Products, Leesville, South Carolina.

The petitioner has requested that the petition be withdrawn. Consequently, the investigation has been terminated.

Signed at Washington, DC, this 24th day of October, 2003.

Richard Church,

*Certifying Officer, Division of Trade
Adjustment Assistance.*

[FR Doc. 03-30083 Filed 12-2-03; 8:45 am]

BILLING CODE 4510-30-P

**NUCLEAR REGULATORY
COMMISSION****[Docket No. 030-34629]****Gamma Knife Center of the Pacific
Environmental Assessment and Final
Finding of No Significant Impact;
Exemption**

The U.S. Nuclear Regulatory Commission is authorizing the Gamma Knife Center of the Pacific, License No. 53-11966-02, an exemption to 10 CFR 35.655, for 120 days to permit the licensee to continue the medical use of its gamma stereotactic radiosurgery unit beyond the interval specified in the regulation for full inspection and servicing.

Environmental Assessment*Identification of the Proposed Action*

The Gamma Knife Center of the Pacific has a United States Nuclear Regulatory Commission (NRC) license (License No. 53-11966-02) that authorizes the medical use of a Leksell Gamma Stereotactic Radiosurgery unit. The licensee has requested, in a letter dated September 8, 2003, that the NRC grant a temporary exemption from the requirements of 10 CFR 35.655(a) in order to continue treating patients until the manufacturer's scheduled source replacement in February 2004. This regulation requires that all stereotactic radiosurgery units be fully inspected and serviced at source replacement or at

intervals not to exceed 5 years whichever comes first.

The licensee had the Leksell Gamma Stereotactic Radiosurgery unit installed in October 1998 and treated the first patient December 1, 1998. The period of time between manufacturer's scheduled source exchange and the last full inspection service before treatment of the first patient exceeds the required 5 year interval. The licensee in e-mails dated September 11 through 18 clarifies that the manufacturer has indicated the semiannual service, which was last performed July 29, 2003 on the licensee's unit, includes all the tests that can be performed without disassembly of the unit. Such disassembly is not advised at any time other than source replacement because a new Leksell Gamma Stereotactic Radiosurgery unit can contain 6,000 curies of cobalt 60 and disassembly can create high radiation fields unless it is performed at source replacement.

Need for the Proposed Action

The exemption is needed so that Gamma Knife Center of the Pacific can continue to provide medical treatment to its patients. The exemption would allow the Gamma Knife Center of the Pacific to use the gamma stereotactic radiosurgery unit until the manufacturer can remove the sources and perform the full inspection and servicing during source replacement. This inspection and service can only be performed after source removal during source replacement. The exemption would permit continued use of the unit and provide needed timely patient therapeutic services without interruption. The source exchange and inspection service is scheduled for February 2004. The 120-day duration of the exemption allows for flexibility if there is a delay in the manufacturer's ability to ship the sources and perform the source exchange. NRC inspections since 1999 have not identified any violations or medical events associated with the use of the gamma stereotactic radiosurgery unit.

Environmental Impacts of the Proposed Action

The gamma stereotactic radiosurgery unit contains 201 cobalt 60 sealed sources and no material will be released into the environment. All the cobalt-60 is contained within the gamma stereotactic radiosurgery unit, as verified by periodic source leak tests performed by the licensee. The proposed action does not increase public radiation exposure. There will be no impact on the environment as a result of the proposed action.