

manufacturer. Section 213 of FDAMA amended section 519(b) of the act. This amendment legislated the replacement of a universal user facility reporting by a system that is limited to a “\* \* \* subset of user facilities that constitutes a representative profile of user reports” for device related deaths and serious injuries. This amendment is reflected in section 519(b)(5)(A) of the act.

FDA is the regulatory agency responsible for the safety and effectiveness of medical products including medical devices and radiological products. Important questions about medical devices, such as those concerning user experience, durability, and rare effects may not be answered until after the device has been marketed. To protect the public health, FDA must be able to rapidly collect information pertaining to adverse events associated with medical devices after they have been marketed. This system is called the Medical Product Surveillance Network (MedSun). The current universal reporting system remains in place during the pilot stages of the new program, and until FDA implements the new national system by regulation. This legislation provides FDA with the

opportunity to design and implement a national surveillance network, composed of well-trained clinical facilities, to provide high quality data on medical devices in clinical use.

Before writing a regulation to implement the large-scale national MedSun reporting system, FDA has been conducting a pilot project to ensure all aspects of the new system address the needs of both the reporting facilities and FDA. This pilot project began with a small sample (approximately 25) and was planned to increase to a larger sample of approximately 250 facilities over a period of approximately 3 years. Data collection began in February 2002 and has been increasing since that time. FDA has achieved its recruitment goals each year, reaching 180 sites at the end of fiscal year (FY) 2003. FDA will reach a total of 240 for FY 2004 and will reach the final goal of 250 by FY 2005. The program has proven to be very popular with sites as FDA has gained a national reputation, with hospitals waiting in line to join.

However, FDA's current resources will not permit FDA to expand beyond 250 sites at this time.

The pilot originally had the following three parts to the data collection: (1) Collecting demographic profile information about the participation facilities, (2) implementing an electronic version of the portions of the MedWatch form (FDA Form No. 3500A, OMB control number 0910-0291) used to report adverse events occurring with medical devices, and (3) adding additional voluntary questions to the data collection. To date, these three features remain unchanged. However, there has been an addition to the data collection that was approved by OMB in the spring of 2004. Therefore, the fourth part of the collection system is the Medical Device Engineering Network (M-DEN)—a place on the MedSun software for the reporters to share information with each other.

In the **Federal Register** of January 27, 2004 (69 FR 3922), FDA published a 60-day notice requesting public comment on the information collection provisions. No comments were received.

FDA estimates the burden of this collection of information as follows:

TABLE 1.—ESTIMATED ANNUAL REPORTING BURDEN<sup>1</sup>

Data Type	No. of Respondents	Annual Frequency per Response	Total Annual Responses	Hours per Response	Total Hours
MedSun <sup>2</sup>	250	8	2,000	.75	1,500
M-DEN <sup>3</sup>	83	10	830	.50	415
Total					1,915

<sup>1</sup> There are no capital costs or operating and maintenance costs associated with this collection.

<sup>2</sup> MedSun means Medical Product Surveillance Network.

<sup>3</sup> M-DEN means Medical Device Engineering Network.

Currently, FDA has 180 sites participating in MedSun pilot program, but expects to have 250 sites over the next 2 years. The frequency of response reflects what FDA has actually been receiving as the average number of submissions in the MedSun Program. While six is the actual average for submissions, FDA hopes to increase this number to eight once their educational materials reach potential respondents. The time estimated to respond is based on feedback FDA has received from current MedSun reporters.

At this time, FDA estimates that one-third of the total number of respondents will access M-DEN aspect of the MedSun software, or approximately 83 persons per year. Each respondent is expected to post 5 problems and respond to 5 problems posted by other MedSun participants for a total of 10

responses per year. It is expected that each visit to the bulletin will not take longer than 30 minutes.

Dated: June 4, 2004.

**Jeffrey Shuren,**

*Assistant Commissioner for Policy.*

[FR Doc. 04-13211 Filed 6-10-04; 8:45 am]

BILLING CODE 4160-01-S

## DEPARTMENT OF HEALTH AND HUMAN SERVICES

### Food and Drug Administration

[Docket No. 2004N-0034]

#### Agency Information Collection Activities; Submission for Office of Management and Budget Review; Comment Request; Medical Devices; Current Good Manufacturing Practice Quality System Regulation

**AGENCY:** Food and Drug Administration, HHS.

**ACTION:** Notice.

**SUMMARY:** The Food and Drug Administration (FDA) is announcing that a proposed collection of information has been submitted to the Office of Management and Budget

(OMB) for review and clearance under the Paperwork Reduction Act of 1995.

**DATES:** Fax written comments on the collection of information by July 14, 2004.

**ADDRESSES:** OMB is still experiencing significant delays in the regular mail, including first class and express mail, and messenger deliveries are not being accepted. To ensure that comments on the information collection are received, OMB recommends that comments be faxed to the Office of Information and Regulatory Affairs, OMB, Attn: Fumie Yokota, Desk Officer for FDA, FAX: 202-395-6974.

**FOR FURTHER INFORMATION CONTACT:** Peggy Robbins, Office of Management Programs (HFA-250), Food and Drug Administration, 5600 Fishers Lane, Rockville, MD 20857, 301-827-1223.

**SUPPLEMENTARY INFORMATION:** In compliance with 44 U.S.C. 3507, FDA has submitted the following proposed collection of information to OMB for review and clearance.

**Medical Devices; Current Good Manufacturing Practice (CGMP) Quality System (QS) Regulations—21 CFR Part 820 (OMB Control Number 0910-0073)—Extension**

Under section 520(f) of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360j(f)), the Secretary of the Department of Health and Human Services (the Secretary) has the authority to prescribe regulations requiring that the methods used in, and the facilities and controls used for, the manufacture, pre-production design validation (including a process to assess the performance of a device but not including an evaluation of the safety and effectiveness of a device), packing, storage, and installation of a device conform to CGMP, as described in such regulations, to assure that the device will be safe and effective and otherwise in compliance with the act.

The CGMP/QS regulation implementing the authority provided by this statutory provision is found at part 820 (21 CFR part 820) and sets forth basic CGMP requirements governing the design, manufacture, packing, labeling, storage, installation, and servicing of all finished medical devices intended for human use. The authority for this regulation is covered under the act (21 U.S.C. 351, 352, 360, 360c, 360d, 360e, 360h, 360i, 360j, 360l, 371, 374, 381, and 383). The CGMP/QS regulation includes requirements for purchasing and service controls, clarifies recordkeeping requirements for device failure and complaint investigations, clarifies requirements for verifying/

validating production processes and process or product changes, and clarifies requirements for product acceptance activities quality data evaluations and corrections of nonconforming product/quality problems. Requirements are compatible with specifications in international quality standards, ISO (International Organization for Standardization) 9001 entitled "Quality Systems Model for Quality Assurance in Design/Development, Production, Installation, and Servicing." CGMP/QS information collections will assist FDA inspections of manufacturer compliance with quality system requirements encompassing design, production, installation, and servicing processes.

Section 820.20(a) through (e) requires management with executive responsibility to establish, maintain, and/or review these topics: The quality policy, the organizational structure, the quality plan, and the quality system procedures of the organization. Section 820.22 requires the conduct and documentation of quality system audits and reaudits. Section 820.25(b) requires the establishment of procedures to identify training needs and documentation of such training.

Section 820.30(a)(1) and (b) through (j) requires, in the following respective order, the establishment, maintenance, and/or documentation of these topics: (1) Procedures to control design of class III and class II devices, and certain class I devices as listed therein; (2) plans for design and development activities and updates; (3) procedures identifying, documenting, and approving design input requirements; (4) procedures defining design output, including acceptance criteria, and documentation of approved records; (5) procedures for formal review of design results and documentation of results in the design history file (DHF); (6) procedures for verifying device design and documentation of results and approvals in the DHF; (7) procedures for validating device design, including documentation of results in the DHF; (8) procedures for translating device design into production specifications; (9) procedures for documenting, verifying validating approved design changes before implementation of changes; and (10) the records and references constituting the DHF for each type of device.

Section 820.40 requires manufacturers to establish and maintain procedures controlling approval and distribution of required documents and document changes.

Section 820.40(a) and (b) requires the establishment and maintenance of

procedures for the review, approval, issuance and documentation of required records (documents) and changes to those records.

Section 820.50(a)(1), (a)(2), (a)(3), and (b) requires the establishment and maintenance of procedures and requirements to ensure service and product quality, records of acceptable suppliers, and purchasing data describing specified requirements for products and services.

Sections 820.60 and 820.65 require, respectively, the establishment and maintenance of procedures for identifying all products from receipt to distribution and for using control numbers to track surgical implants and life-sustaining or supporting devices and their components.

Section 820.70(a)(1) through (a)(5), (b) through (e), (g)(1) through (g)(3), and (h) and (i) requires the establishment, maintenance, and/or documentation of these topics: (1) Process control procedures; (2) procedures for verifying or validating changes to specification, method, process, or procedure; (3) procedures to control environmental conditions and inspection result records; (4) requirements for personnel hygiene; (5) procedures for preventing contamination of equipment and products; (6) equipment adjustment, cleaning and maintenance schedules; (7) equipment inspection records; (8) equipment tolerance postings; procedures for utilizing manufacturing materials expected to have an adverse effect on product quality; and (9) validation protocols and validation records for computer software and software changes.

Sections 820.72(a) and (b)(1) and (b)(2) and 820.75(a) through (c) require, respectively, the establishment, maintenance, and/or documentation of these topics: (1) Equipment calibration and inspection procedures; (2) national, international or in-house calibration standards; (3) records that identify calibrated equipment and next calibration dates; (4) validation procedures and validation results for processes not verifiable by inspections and tests; (5) procedures for keeping validated processes within specified limits; (6) records for monitoring and controlling validated processes; and (7) records of the results of revalidation where necessitated by process changes or deviations.

Sections 820.80(a) through (e) and 820.86, respectively, require the establishment, maintenance, and/or documentation of these topics: (1) Procedures for incoming acceptance by inspection, test or other verification; (2) procedures for ensuring that in-process

products meet specified requirements and the control of product until inspection and tests are completed; (3) procedures for, and records that show, incoming acceptance or rejection is conducted by inspections, tests or other verifications; (4) procedures for, and records that show, finished devices meet acceptance criteria and are not distributed until device master record (DMR) activities are completed; (5) records in the device history record (DHR) showing acceptance dates, results and equipment used; and (6) the acceptance/rejection identification of products from receipt to installation and servicing.

Sections 820.90(a), (b)(1), (b)(2), and 820.100 require, respectively, the establishment, maintenance and/or documentation of these topics: (1) Procedures for identifying, recording, evaluating and disposing of nonconforming product; (2) procedures for reviewing and recording concessions made for, and disposition of, nonconforming product; (3) procedures for reworking products, evaluating possible adverse rework effect and recording results in the DHR; (4) procedures and requirements for corrective and preventive actions, including analysis, investigation, identification and review of data, records, causes and results; and (5) records for all corrective and preventive action activities.

Section 820.100(a)(1) through (a)(7) states that procedures and requirements shall be established and maintained for corrective/preventive actions, including the following: (1) Analysis of data from process, work, quality, servicing records; investigation of nonconformance causes; (2) identification of corrections and their effectiveness; (3) recording of changes made; and, (4) appropriate distribution and managerial review of corrective and preventive action information.

Section 820.120 states that manufacturers shall establish/maintain procedures to control labeling storage/application; and examination/release for storage and use, and document those procedures.

Sections 820.120(b) and (d), 820.130, 820.140, 820.150(a) and (b), 820.160(a) and (b), and 820.170(a) and (b), respectively, require the establishment, maintenance, and/or documentation of these topics: (1) Procedures for controlling and recording the storage, examination, release and use of labeling; (2) the filing of labels/labeling used in the DHR; (3) procedures for controlling product storage areas and receipt/dispatch authorizations; (4) procedures controlling the release of products for

distribution; (5) distribution records that identify consignee, product, date and control numbers; and (6) instructions, inspection and test procedures that are made available, and the recording of results for devices requiring installation.

Sections 820.180(b) and (c), 820.181(a) through (e), 820.184(a) through (f), and 820.186 require, respectively, the maintenance of records: (1) That are retained at prescribed site(s), made readily available and accessible to FDA and retained for the device's life expectancy or for 2 years; (2) that are contained or referenced in a DMR consisting of device, process, quality assurance, packaging and labeling, and installation, maintenance, and servicing specifications and procedures; (3) that are contained in DHRs, demonstrate the manufacture of each unit, lot or batch of product in conformance with DMR and regulatory requirements, and include manufacturing and distribution dates and quantities, acceptance documents, labels and labeling, and control numbers; and (4) that are contained in a quality system record (QSR) consisting of references, documents, procedures and activities not specific to particular devices.

Sections 820.198(a) through (c) and 820.200(a) and (d), respectively, require the establishment, maintenance and/or documentation of these topics: (1) Complaint files and procedures for receiving, reviewing and evaluating complaints; (2) complaint investigation records identifying the device, complainant and relationship of the device to the incident; (3) complaint records that are reasonably accessible to the manufacturing site or at prescribed sites; (4) procedures for performing and verifying that device servicing requirements are met and that service reports involving complaints are processed as complaints; and (5) service reports that record the device, service activity, and test and inspection data.

Section 820.250 requires the establishment and maintenance of procedures to identify valid statistical techniques necessary to verify process and product acceptability; and sampling plans, when used, that are written and based on a valid statistical rationale, and procedures for ensuring adequate sampling methods.

The CGMP/QS regulation amends and revises the CGMP requirements for medical devices set out at part 820. It adds design and purchasing controls; modifies previous critical device requirements; revises previous validation and other requirements; and harmonizes device CGMP requirements with quality system specifications in the

international standard, ISO 9001:1994 entitled "Quality Systems—Model for Quality Assurance in Design, Development Production, Installation and Servicing." The regulation applies neither to manufacturers of components or parts of finished devices, nor to manufacturers of human blood and blood components subject to 21 CFR part 606. With respect to devices classified in class I, design control requirements apply only to class I devices listed in § 820.30(a)(2).

The regulation imposes burdens upon finished device manufacturer firms, which are subject to all recordkeeping requirements, and also upon finished device contract manufacturer, specification developer, repacker and relabeler, and contract sterilizer firms, which are subject only to requirements applicable to their activities. Due to modifications to the guidance given for remanufacturers of hospital single-use devices, reusers of hospital single-use devices will now be considered to have the same requirements as manufacturers in regard to this regulation. The establishment, maintenance and/or documentation of procedures, records and data required by this regulation will assist FDA in determining whether firms are in compliance with CGMP requirements, which are intended to ensure that devices meet their design, production, labeling, installation, and servicing specifications and, thus are safe, effective and suitable for their intended purpose. In particular, compliance with CGMP design control requirements should decrease the number of design-related device failures that have resulted in deaths and serious injuries.

If FDA did not impose these recordkeeping requirements, it anticipates that design-related device failures would continue to occur in the same numbers as before and continue to result in a significant number of device recalls and preventable deaths and serious injuries. Moreover, manufacturers would be unable to take advantage of substantial savings attributable to reduced recall costs, improved manufacturing efficiency, and improved access to international markets through compliance with CGMP requirements that are harmonized with international quality system standards.

The CGMP/QS regulation applies to some 8,254 respondents. These recordkeepers consist of 8,188 original respondents and an estimated 66 hospitals which remanufacture or reuse single use medical devices. They include manufacturers, subject to all requirements and contract

manufacturers, specification developers, repackers/relabelers and contract sterilizers, subject only to requirements applicable to their activities. Hospital remanufacturers of single use medical devices (SUDs) are now defined to be manufacturers under guidelines issued by FDA's Center for Devices and Radiological Health's (CDRH) Office of Surveillance and Biometrics. Respondents to this collection have no reporting activities, but must make required records available for review or

copying during FDA inspection. The regulation contains additional recordkeeping requirements in such areas as design control, purchasing, installation, and information relating to the remanufacture of single use medical devices. The estimates for burden are derived from those incremental tasks that were determined when the new CGMP/QS regulation became final as well as those carry-over requirements. The carry-over requirements are based on decisions made by the agency on July

16, 1992, under OMB control number 0910-0073. This still provides valid baseline data.

FDA estimates respondents will have a total annual recordkeeping burden of approximately 2,833,020 hours. This figure also consists of approximately 143,052 hours spent on a startup basis by 650 new firms. FDA estimates information collection burdens imposed as follows:

TABLE 1.—ESTIMATED ANNUAL RECORDKEEPING BURDEN<sup>1</sup>

CFR Section	Number of Recordkeepers	Annual Frequency of Recordkeeping	Total Annual Hours	Hours per Recordkeeper	Total Hours	Total Operating and Maintenance Cost
820.20(a)	8,254	1	8,254	6.58	54,311	-----
820.20(b)	8,254	1	8,254	4.43	36,565	-----
820.20(c)	8,254	1	8,254	6.17	50,927	-----
820.20(d)	8,254	1	8,254	9.89	81,632	-----
820.20(e)	8,254	1	8,254	9.89	81,632	-----
820.22	8,254	1	8,254	32.72	270,071	-----
820.25(b)	8,254	1	8,254	12.68	104,661	-----
820.30(a)(1)	8,254	1	8,254	1.75	14,445	-----
820.30(b)	8,254	1	8,254	5.95	49,111	-----
820.30(c)	8,254	1	8,254	1.75	14,445	-----
820.30(d)	8,254	1	8,254	1.75	14,445	-----
820.30(e)	8,254	1	8,254	23.39	193,061	-----
820.30(f)	8,254	1	8,254	37.42	308,865	-----
820.30(g)	8,254	1	8,254	37.42	308,865	-----
820.30(h)	8,254	1	8,254	3.34	27,568	-----
820.30(i)	8,254	1	8,254	17.26	142,464	-----
820.30(j)	8,254	1	8,254	2.64	21,791	-----
820.4	8,254	1	8,254	8.91	73,543	-----
820.40(a)-(b)	8,254	1	8,254	2.04	16,838	-----
820.50(a)(1)-(a)(3)	8,254	1	8,254	21.9	180,763	\$1,181,925
820.50(b)	8,254	1	8,254	6.02	49,689	-----
820.60	8,254	1	8,254	0.32	2,641	-----
820.65	8,254	1	8,254	0.67	5,530	-----
820.70(a)(1)-(a)(5)	8,254	1	8,254	1.85	15,270	-----
820.70(b)-(c)	8,254	1	8,254	1.85	15,270	-----
820.70(d)	8,254	1	8,254	2.87	23,689	-----
820.70(e)	8,254	1	8,254	1.85	15,270	-----

TABLE 1.—ESTIMATED ANNUAL RECORDKEEPING BURDEN<sup>1</sup>—Continued

CFR Section	Number of Recordkeepers	Annual Frequency of Recordkeeping	Total Annual Hours	Hours per Recordkeeper	Total Hours	Total Operating and Maintenance Cost
820.70(g)(1)-(g)(3)	8,254	1	8,254	1.43	11,803	-----
820.70(h)	8,254	1	8,254	1.85	15,270	-----
820.72(a)	8,254	1	8,254	4.92	40,610	-----
820.70(i)	8,254	1	8,254	7.5	61,905	-----
820.72(b)(1) to (b)(2)	8,254	1	8,254	1.43	11,803	-----
820.75(a)	8,254	1	8,254	2.69	22,203	-----
820.75(b)	8,254	1	8,254	1.02	8,419	-----
820.75(c)	8,254	1	8,254	1.11	9,162	-----
820.80(a)-(e)	8,254	1	8,254	4.8	39,619	-----
820.86	8,254	1	8,254	0.79	6,521	-----
820.90(a)	8,254	1	8,254	4.95	40,857	-----
820.90(b)(1)-(b)(2)	8,254	1	8,254	4.95	40,857	-----
820.100(a)(1)-(a)(7)	8,254	1	8,254	12.48	103,010	-----
820.100(b)	8,254	1	8,254	1.28	10,565	-----
820.120	8,254	1	8,254	0.45	3,714	-----
820.120(b)	8,254	1	8,254	0.45	3,714	-----
820.120(d)	8,254	1	8,254	0.45	3,714	-----
820.130	8,254	1	8,254	0.45	3,714	-----
820.140	8,254	1	8,254	6.34	52,330	-----
820.150(a)-(b)	8,254	1	8,254	5.67	46,800	-----
820.160(a)-(b)	8,254	1	8,254	0.67	5,530	-----
820.170(a)-(b)	8,254	1	8,254	1.5	12,381	-----
820.180(b)-(c)	8,254	1	8,254	1.5	12,381	-----
820.181((a)-(e)	8,254	1	8,254	1.21	9,987	-----
820.184(a)-(f)	8,254	1	8,254	1.41	11,638	-----
820.186	8,254	1	8,254	0.4	3,302	-----
820.198(a)-(c)	8,254	1	8,254	4.94	40,775	-----
820.200(a) and (d)	8,254	1	8,254	2.61	21,543	-----
820.250	8,254	1	8,254	0.67	5,530	-----
Totals					2,283,020	\$1,181,925

<sup>1</sup> There are no capital costs or operating and maintenance costs associated with this collection of information.

Burden (labor) hour and cost estimates were originally developed under FDA contract by Eastern Research

Group, Inc. (ERG), in 1996 when the CGMP/QS regulation became final. These figures are still accurate.

Additional factors considered in deriving estimates included:

• Establishment type: Query has been made of CDRH's registration/listing databank and has counted 8,188 domestic firms subject to CGMPs. In addition, hospitals which reuse or remanufacture devices are now considered manufacturers under new FDA guidance. During the last report, it was estimated that out of the 6,000 hospitals in the United States, one third of them (or 2,000 hospitals) will reuse or remanufacture single use medical devices. After investigations of many hospitals and the changes in enforcements of FDA's requirements for hospitals, the number of reuse or remanufactures of single-use medical devices have decreased from the estimated 2,000 to an estimated 66 hospitals. Thus, the number of manufacturers will increase from 7,229 to 8,188, but the total number of firms subject to CGMPs will decrease from 9,229 to 8,254.

• Potentially affected establishments: Except for manufacturers, not every type of firm is subject to every CGMP/QS requirement. For example, all are subject to FDA's quality policy regulations (§ 820.20(a)), document control regulations (§ 820.40), and other requirements, whereas only manufacturers and specification developers are subject to FDA's design controls regulations (§ 820.30). The type of firm subject to each requirement was identified by ERG.

FDA estimated the burden hours (and costs) for the previous CGMP regulation in 1992. That estimate was submitted to OMB on May 4, 1992. It was approved by OMB on July 16, 1992, and expired on June 30, 1995. The methodology used is different than that used by ERG in estimating incremental tasks when the new CGMP/QS became final. Nevertheless, the agency believes its 1992 estimate adequately represents labor hours (and costs) needed to comply with previous CGMP requirements carried over into the new CGMP/QS regulation. The 1992 estimate used 9,289 respondents (rather than 8,254 respondents), which compensates for differences in methodology.

FDA estimates that some 650 "new" establishments (marketing devices for the first time) will expend some 143,052 "development" hours on a one-time startup basis to develop records and procedures for the CGMP/QS regulation.

FDA estimates that annual labor hours are apportioned as follows: 40 percent goes to requirements dealing with manufacturing specifications, process controls and the DHR; 20 percent goes to requirements dealing with components and acceptance activities; 25 percent goes to requirements dealing

with equipment, records (the DMR and QSR), complaint investigations, labeling/packaging and reprocessing/ investigating product nonconformance; and 15 percent goes to quality audit, traceability, handling, distribution, statistical, and other requirements.

Dated: June 4, 2004.

**Jeffrey Shuren,**

*Assistant Commissioner for Policy.*

[FR Doc. 04-13212 Filed 6-10-04; 8:45 am]

**BILLING CODE 4160-01-S**

## DEPARTMENT OF HEALTH AND HUMAN SERVICES

### Food and Drug Administration

[Docket No. 2003N-0425]

#### **Agency Information Collection Activities; Announcement of Office of Management and Budget Approval; Substances Prohibited From Use in Animal Food or Feed; Animal Proteins Prohibited in Ruminant Feed**

**AGENCY:** Food and Drug Administration, HHS.

**ACTION:** Notice.

**SUMMARY:** The Food and Drug Administration (FDA) is announcing that a collection of information entitled "Substances Prohibited From Use in Animal Food or Feed; Animal Proteins Prohibited in Ruminant Feed" has been approved by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (the PRA).

**FOR FURTHER INFORMATION CONTACT:** Denver Presley, Office of Management Programs (HFA-250), Food and Drug Administration, 5600 Fishers Lane, Rockville, MD 20857, 301-827-1472.

**SUPPLEMENTARY INFORMATION:** In the **Federal Register** of January 16, 2004 (69 FR 2602), the agency announced that the proposed information collection had been submitted to OMB for review and clearance under 44 U.S.C. 3507. An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has now approved the information collection and has assigned OMB control number 0910-0339. The approval expires on May 31, 2007. A copy of the supporting statement for this information collection is available on the Internet at <http://www.fda.gov/ohrms/dockets>.

Dated: June 4, 2004.

**Jeffrey Shuren,**

*Assistant Commissioner for Policy.*

[FR Doc. 04-13213 Filed 6-10-04; 8:45 am]

**BILLING CODE 4160-01-S**

## DEPARTMENT OF HEALTH AND HUMAN SERVICES

### Food and Drug Administration

[Docket No. 2003N-0542]

#### **Agency Information Collection Activities; Announcement of OMB Approval; Premarket Notification Submissions**

**AGENCY:** Food and Drug Administration, HHS.

**ACTION:** Notice.

**SUMMARY:** The Food and Drug Administration (FDA) is announcing that a collection of information entitled "Premarket Notification Submissions" has been approved by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995.

**FOR FURTHER INFORMATION CONTACT:** Peggy Robbins, Office of Management Programs (HFA-250), Food and Drug Administration, 5600 Fishers Lane, Rockville, MD 20857, 301-827-1223.

**SUPPLEMENTARY INFORMATION:** In the **Federal Register** of March 9, 2004 (69 FR 11022), the agency announced that the proposed information collection had been submitted to OMB for review and clearance under 44 U.S.C. 3507. An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has now approved the information collection and has assigned OMB control number 0910-0120. The approval expires on May 31, 2007. A copy of the supporting statement for this information collection is available on the Internet at <http://www.fda.gov/ohrms/dockets>.

Dated: June 4, 2004.

**Jeffrey Shuren,**

*Assistant Commissioner for Policy.*

[FR Doc. 04-13214 Filed 6-10-04; 8:45 am]

**BILLING CODE 4160-01-S**

## DEPARTMENT OF HEALTH AND HUMAN SERVICES

### Food and Drug Administration

[Docket No. 2004N-0245]

#### **Agency Information Collection Activities; Proposed Collection; Comment Request; Current Good Manufacturing Practice Regulations for Medicated Feeds**

**AGENCY:** Food and Drug Administration, HHS.

**ACTION:** Notice.