

(i) The clinical data must be representative of the intended use population for the device. Any selection criteria or sample limitations must be fully described and justified.

(ii) The assessment must demonstrate output consistency using the expected range of data sources and data quality encountered in the intended use population and environment.

(iii) The assessment must compare device output with a clinically accurate patient-contacting relevant comparator device in an accurate and reproducible manner.

(4) A human factors and usability engineering assessment must be provided that evaluates the risk of improper measurement.

(5) Labeling must include:

(i) A description of what the device measures and outputs to the user;

(ii) Warnings identifying sensor acquisition factors or subject conditions or characteristics (garment types/textures, motion, etc.) that may impact measurement results;

(iii) Guidance for interpretation of the measurements, including a statement that the output is adjunctive to other physical vital sign parameters and patient information;

(iv) The expected performance of the device for all intended use populations and environments; and

(v) Robust instructions to ensure correct system setup.

Dated: January 3, 2023.

Lauren K. Roth,

Associate Commissioner for Policy.

[FR Doc. 2023–00010 Filed 1–5–23; 8:45 am]

BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 874

[Docket No. FDA–2022–N–3171]

Medical Devices; Ear, Nose, and Throat Devices; Classification of the Powered Insertion System for a Cochlear Implant Electrode Array

AGENCY: Food and Drug Administration, Department of Health and Human Services (HHS).

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA, Agency, or we) is classifying the powered insertion system for a cochlear implant electrode array into class II (special controls). The special controls that apply to the device

type are identified in this order and will be part of the codified language for the powered insertion system for a cochlear implant electrode array's classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices.

DATES: This order is effective January 6, 2023. The classification was applicable on October 1, 2021.

FOR FURTHER INFORMATION CONTACT:

Vasant Dasika, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 1206, Silver Spring, MD, 20993–0002, 301–796–5365, Vasant.Dasika@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA has classified the powered insertion system for a cochlear implant electrode array as class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified as, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (see 21 U.S.C. 360c(f)(1)). We refer to these devices as “postamendments devices” because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (see 21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through “De Novo” classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act. Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105–115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112–144) modified the De Novo application process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a 510(k) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining “substantial equivalence”). Instead, sponsors can use the less-burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On December 18, 2019, FDA received *iotaMotion, Inc.*'s request for De Novo classification of the *iotaSOFT* Insertion System—Drive Unit, Controller and Accessories. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable

assurance of safety and effectiveness, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see 21 U.S.C. 360c(a)(1)(B)). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA

has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on October 1, 2021, FDA issued an order to the requester classifying the device into class II. In this final order, FDA is codifying the classification of the device by adding 21 CFR 874.4450.¹ We have named the generic type of device powered

insertion system for a cochlear implant electrode array, and it is identified as a prescription device used to assist in placing an electrode array into the cochlea.

FDA has identified the following risks to health associated specifically with this type of device and the measures required to mitigate these risks in table 1.

TABLE 1—POWERED INSERTION SYSTEM FOR A COCHLEAR IMPLANT ELECTRODE RISKS AND MITIGATION MEASURES

Identified risks	Mitigation measures
Risks to health relating to device interface with patient anatomy, including: <ul style="list-style-type: none">• Damage to skull tissue.• Damage to dura mater.• Bone damage.• Cerebrospinal fluid leak.• Damage to cochlea; hearing loss, tinnitus, vertigo.	Clinical performance testing, Usability testing, Non-clinical performance testing, and Labeling.
Cochlear implant insertion failure leading to: <ul style="list-style-type: none">• Trauma to cochlear structures resulting in residual hearing loss or nerve degeneration.• Suboptimal array placement (including array rotation) leading to poor hearing performance.• Failure to disengage from cochlear implant at end of procedure, leading to manual correction and insertion.	Clinical performance testing, Non-clinical performance testing, Usability testing, Cochlear implant compatibility validation, Software verification, validation, and hazard analysis, and Labeling.
Damage to cochlear implant during insertion leading to poor cochlear implant performance and/or compromised implant reliability.	Non-clinical performance testing, Usability testing, Cochlear implant compatibility validation, Shelf life testing, Software verification, validation, and hazard analysis, and Labeling.
Adverse tissue reaction, including irritation/inflammation of surgical site	Biocompatibility evaluation.
Electromagnetic interference, thermal injury, or electric shock	Electrical safety testing, Electromagnetic compatibility testing, and Labeling.
Infection	Sterilization validation, Shelf life testing, and Labeling.
Excessive operation time leading to increased exposure to anesthesia	Clinical performance testing, Usability testing, and Labeling.

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this order. This device is subject to premarket notification requirements under section 510(k) of the FD&C Act.

At the time of classification, powered insertion systems for a cochlear implant electrode array are for prescription use only. Prescription devices are exempt from the requirement for adequate directions for use for the layperson under section 502(f)(1) of the FD&C Act (21 U.S.C. 352(f)(1)) and 21 CFR 801.5, as long as the conditions of 21 CFR 801.109 are met.

¹ FDA notes that the “ACTION” caption for this final order is styled as “Final amendment; final order,” rather than “Final order.” Beginning in December 2019, this editorial change was made to

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3521). The collections of information in 21 CFR part 860, subpart D, regarding De Novo classification have been approved under OMB control number 0910–0844; the collections of information in 21 CFR

indicate that the document “amends” the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register’s (OFR) interpretations of the Federal Register Act (44

part 814, subparts A through E, regarding premarket approval, have been approved under OMB control number 0910–0231; the collections of information in part 807, subpart E, regarding premarket notification submissions, have been approved under OMB control number 0910–0120; the collections of information in 21 CFR part 820, regarding quality system regulation, have been approved under OMB control number 0910–0073; and the collections of information in 21 CFR part 801, regarding labeling, have been approved under OMB control number 0910–0485.

List of Subjects in 21 CFR Part 874

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 874 is amended as follows:

U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

PART 874—EAR, NOSE, AND THROAT DEVICES

■ 1. The authority citation for part 874 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

■ 2. Add § 874.4450 to subpart E to read as follows:

§ 874.4450 Powered insertion system for a cochlear implant electrode array.

(a) *Identification.* A powered insertion system for a cochlear implant electrode array is a prescription device used to assist in placing an electrode array into the cochlea.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use, including evaluation of all adverse events.

(2) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use. Testing must include:

(i) Verification of cochlear implant attachment force, release force, and insertion speed;

(ii) Testing to demonstrate the device does not damage or degrade the cochlear implant (including the lead and array portions of the cochlear implant); and

(iii) Comparison testing with manual insertion to evaluate:

(A) Differences in cochlear implant array insertion force associated with use of the device; and

(B) Intracochlear placement of the cochlear implant array (intended scala placement and array insertion depth, together with minimal array tip foldover and cochlear scala translocation).

(3) Usability testing in a simulated hospital environment with an anatomically relevant model (*e.g.*, cadaver testing) that evaluates the following:

(i) Successful use to aid in placement of the electrode array into the cochlea; and

(ii) Harms caused by use errors observed.

(4) Changes in cochlear implant compatibility are determined to significantly affect the safety or effectiveness of the device and must be validated through performance testing or a rationale for omission of any testing.

(5) The patient-contacting components of the device must be demonstrated to be biocompatible.

(6) Performance testing must demonstrate the electromagnetic

compatibility, electrical safety, and thermal safety of the device.

(7) The patient-contacting components of the device must be demonstrated to be sterile and non-pyrogenic.

(8) Performance testing must support the shelf life of device components provided sterile by demonstrating continued sterility, package integrity, and device functionality over the labeled shelf life.

(9) Software verification, validation, and hazard analysis must be performed for any software components of the device.

(10) Labeling must include:

(i) The recommended training for the safe use of the device;

(ii) Summary of the relevant clinical and non-clinical testing pertinent to use of the device with compatible electrode arrays; and

(iii) A shelf life.

Dated: January 3, 2023.

Lauren K. Roth,

Associate Commissioner for Policy.

[FR Doc. 2023-00008 Filed 1-5-23; 8:45 am]

BILLING CODE 4164-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES**Food and Drug Administration****21 CFR Part 888**

[Docket No. FDA-2022-N-3190]

Medical Devices; Orthopedic Devices; Classification of the Resorbable Shoulder Spacer

AGENCY: Food and Drug Administration, Department of Health and Human Services (HHS).

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA, Agency, or we) is classifying the resorbable shoulder spacer into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for the resorbable shoulder spacer's classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices.

DATES: This order is effective January 6, 2023. The classification was applicable on July 12, 2021.

FOR FURTHER INFORMATION CONTACT: Farzana Sharmin, Center for Devices

and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 4564, Silver Spring, MD 20993-0002, 301-796-4067, Farzana.Sharmin@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:**I. Background**

Upon request, FDA has classified the resorbable shoulder spacer as class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified as, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (see 21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (see 21 U.S.C. 360c(ii)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act. Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo application process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a 510(k) for a device that has not previously been classified. After receiving an order from FDA classifying