

operation, and maintenance instructions for the VSCC™ as well as a new equivalent method identification label to be installed on the sampler. If the conversion is to be permanent, the original designated reference method label should be removed from the sampler and replaced with the new designated equivalent method label. In a case where a converted sampler may need to be restored later to its original reference method configuration (such as for an application specifically requiring a reference method) by re-installation of the WINS impactor (or PM₁₀ extension tube), the new equivalent method label may be installed on the sampler without removing the original reference method label, such that the sampler bears both labels. (Alternatively, the new label may describe multiple configurations.) In this situation, the sampler shall be clearly and conspicuously marked by the operator to indicate its current configuration (*i.e.* WINS/PM_{2.5} reference method, VSCC™/PM_{2.5} equivalent method, or PM₁₀ reference method) so that the monitoring method is correctly identified and the correct method code is used when reporting monitoring data obtained with the sampler.

Part 53 requires that sellers of designated reference or equivalent method analyzers or samplers comply with certain conditions. These conditions are specified in 40 CFR 53.9 and are summarized below:

(a) A copy of the approved operation or instruction manual must accompany the sampler or analyzer when it is delivered to the ultimate purchaser.

(b) The sampler or analyzer must not generate any unreasonable hazard to operators or to the environment.

(c) The sampler or analyzer must function within the limits of the applicable performance specifications given in 40 CFR parts 50 and 53 for at least one year after delivery when maintained and operated in accordance with the operation or instruction manual.

(d) Any sampler or analyzer offered for sale as part of a reference or equivalent method must bear a label or sticker indicating that it has been designated as part of a reference or equivalent method in accordance with part 53 and showing its designated method identification number.

(e) If such an analyzer has two or more selectable ranges, the label or sticker must be placed in close proximity to the range selector and indicate which range or ranges have been included in the reference or equivalent method designation.

(f) An applicant who offers samplers or analyzers for sale as part of a

reference or equivalent method is required to maintain a list of ultimate purchasers of such samplers or analyzers and to notify them within 30 days if a reference or equivalent method designation applicable to the method has been canceled or if adjustment of the sampler or analyzer is necessary under 40 CFR 53.11(b) to avoid a cancellation.

(g) An applicant who modifies a sampler or analyzer previously designated as part of a reference or equivalent method is not permitted to sell the sampler or analyzer (as modified) as part of a reference or equivalent method (although it may be sold without such representation), nor to attach a designation label or sticker to the sampler or analyzer (as modified) under the provisions described above, until the applicant has received notice under 40 CFR 53.14(c) that the original designation or a new designation applies to the method as modified, or until the applicant has applied for and received notice under 40 CFR 53.8(b) of a new reference or equivalent method determination for the sampler or analyzer as modified.

(h) An applicant who offers PM_{2.5} samplers for sale as part of a reference or equivalent method is required to maintain the manufacturing facility in which the sampler is manufactured as an ISO 9001-certified facility.

(i) An applicant who offers PM_{2.5} samplers for sale as part of a reference or equivalent method is required to submit annually a properly completed Product Manufacturing Checklist, as specified in part 53.

Aside from occasional breakdowns or malfunctions, consistent or repeated noncompliance with any of these conditions should be reported to: Director, Human Exposure and Atmospheric Sciences Division (MD-77), National Exposure Research Laboratory, U.S. Environmental Protection Agency, Research Triangle Park, North Carolina 27711.

Designation of these reference and equivalent methods is intended to assist the States in establishing and operating their air quality surveillance systems under 40 CFR part 58.

Questions concerning the commercial availability or technical aspects of any of these methods should be directed to the appropriate applicant.

Dated: March 21, 2002.

Jewel F. Morris,

Acting Deputy Director, National Exposure Research Laboratory.

[FR Doc. 02-7944 Filed 4-1-02; 8:45 am]

BILLING CODE 6560-50-P

ENVIRONMENTAL PROTECTION AGENCY

[FRL-7167-1]

Environmental Laboratory Advisory Board (ELAB) Meeting Dates, and Agenda

AGENCY: Environmental Protection Agency.

ACTION: Notice of Teleconference Meeting.

SUMMARY: The Environmental Protection Agency's Environmental Laboratory Advisory Board (ELAB) will have a teleconference meeting on April 17, 2002, at 11 A.M. EST to discuss the ideas and views presented at the previous ELAB meetings, as well as new business. Items to be discussed include (1) Review of NELAC mission, (2) update on recommendations to restructure the National Environmental Laboratory Accreditation Conference (NELAC) to allow it to better serve the future needs of EPA, the States, and the private sector, (3) approaches to facilitate NELAP accreditation of smaller environmental laboratories, and (4) Discussion of ELAB recommendations to EPA. ELAB is soliciting input from the public on these and other issues related to the National Environmental Laboratory Accreditation Program (NELAP) and the NELAC standards. Written comments on NELAP laboratory accreditation and the NELAC standards are encouraged and should be sent to Mr. Edward Kantor, DFO, PO Box 93478, Las Vegas NV 89193, faxed to (702) 798-2261, or emailed to kantor.edward@epa.gov. Members of the public are invited to listen to the teleconference calls and, time permitting, will be allowed to comment on issues discussed during this and previous ELAB meetings. Those persons interested in attending should call Edward Kantor at 702-798-2690 to obtain teleconference information. The number of lines are limited and will be distributed on a first come, first serve basis. Preference will be given to a group wishing to attend over a request from an individual.

John G. Lyon

Director, Environmental Sciences Division, National Environmental Research Laboratory.

[FR Doc. 02-7941 Filed 4-1-02; 8:45 am]

BILLING CODE 6560-50-M