

extension of an existing collection of information, and to allow 60 days for public comment in response to the notice. This notice solicits comments on the Proposed Extension without Change and the information collection requirements related to the National Survey of Older Americans Act Participants.

DATES: Comments on the collection of information must be submitted electronically by 11:59 p.m. (EST) or postmarked by October 2, 2023.

ADDRESSES: Submit electronic comments on the collection of information to Kristen.Robinson@acl.hhs.gov. Submit written comments on the collection of information to Administration for Community Living, Washington, DC 20201, Attention: Kristen Robinson.

FOR FURTHER INFORMATION CONTACT: Kristen Robinson, Administration for Community Living, Washington, DC 20201, by email at Kristen.Robinson@acl.hhs.gov, or by telephone at 202-795-7428.

SUPPLEMENTARY INFORMATION: Under the PRA (44 U.S.C. 3501-3520), Federal agencies must obtain approval from the Office of Management and Budget (OMB) for each collection of information they conduct or sponsor. "Collection of information" is defined in 44 U.S.C. 3502(3) and 5 CFR 1320.3(c) and includes agency requests or requirements that members of the public submit reports, keep records, or

provide information to a third party. Section 3506(c)(2)(A) of the PRA (44 U.S.C. 3506(c)(2)(A)) requires Federal agencies to provide a 60-day notice in the **Federal Register** concerning each proposed collection of information, including each proposed extension of an existing collection of information, before submitting the collection to OMB for approval. To comply with this requirement, ACL is publishing a notice of the proposed collection of information set forth in this document. With respect to the following collection of information, ACL invites comments on our burden estimates or any other aspect of this collection of information, including:

(1) whether the proposed collection of information is necessary for the proper performance of ACL's functions, including whether the information will have practical utility;

(2) the accuracy of ACL's estimate of the burden of the proposed collection of information, including the validity of the methodology and assumptions used to determine burden estimates;

(3) ways to enhance the quality, utility, and clarity of the information to be collected; and

(4) ways to minimize the burden of the collection of information on respondents, including using automated collection techniques when appropriate, and other forms of information technology. The National Survey of Older Americans Act (OAA) Participants information collection

includes consumer assessment surveys for the Congregate and Home-delivered meal nutrition programs; Case Management, Homemaker, and Transportation Services; and the National Family Caregiver Support Program. This survey builds on earlier national pilot studies and surveys, as well as performance measurement tools developed by ACL grantees in the Performance Outcomes Measures Project (POMP). Changes identified as a result of these initiatives were incorporated into the last data collection package that was approved by OMB and are included in this proposed extension of a currently approved collection.

This information will be used by ACL to track performance outcome measures; support budget requests; comply with the GPRMA Modernization Act of 2010 (GPRMA) reporting requirements; provide national benchmark information; and inform program development and management initiatives.

Copies of the survey instruments and data from previous National Surveys of OAA Participants can be found and queried using the Aging, Independence, and Disability (AGID) Program Data Portal at <https://agid.acl.gov/>. The proposed data collection tools may be found on the ACL website at <https://www.acl.gov/about-acl/public-input>.

Estimated Program Burden: ACL estimates the annual burden associated with this collection of information as follows:

| Respondent/data collection activity | Number of respondents | Responses per respondent | Hours per response | Annual burden hours |
|---|-----------------------|--------------------------|--------------------|---------------------|
| Area Agency on Aging: Respondent selection process | 300 | 1 | 4.0 | 1,200 |
| Service recipients (i.e., Congregate and Home-Delivered Meal nutrition programs, Case Management, Homemaker, Transportation services) + Rotating Module | 4,000 | 1 | 0.75 | 3,000 |
| National Family Caregiver Support Program clients + Rotating Module | 2,000 | 1 | 0.75 | 1,500 |
| Total | 6,300 | 1 | * 0.90 | 5,700 |

*(weighted mean).

Dated: July 28, 2023.

Alison Barkoff,

Acting Administrator and Assistant Secretary for Aging.

[FR Doc. 2023-16419 Filed 8-1-23; 8:45 am]

BILLING CODE 4154-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA-2004-N-0451]

Food and Drug Administration Modernization Act of 1997: Modifications to the List of Recognized Standards, Recognition List Number: 059

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA or Agency) is announcing a publication containing modifications the Agency is making to the list of standards FDA recognizes for use in premarket reviews (FDA Recognized Consensus Standards). This publication, entitled "Modifications to the List of Recognized Standards, Recognition List Number: 059" (Recognition List Number: 059), will assist manufacturers who elect to declare conformity with consensus standards to meet certain requirements for medical devices.

DATES: Submit either electronic or written comments on the notice at any time. These modifications to the list of recognized standards are applicable August 2, 2023.

ADDRESSES: You may submit comments on the current list of FDA Recognized Consensus Standards at any time as follows:

Electronic Submissions

Submit electronic comments in the following way:

- *Federal eRulemaking Portal:* <https://www.regulations.gov>. Follow the instructions for submitting comments. Comments submitted electronically, including attachments, to <https://www.regulations.gov> will be posted to the docket unchanged. Because your comment will be made public, you are solely responsible for ensuring that your comment does not include any confidential information that you or a third party may not wish to be posted, such as medical information, your or anyone else's Social Security number, or confidential business information, such as a manufacturing process. Please note that if you include your name, contact information, or other information that identifies you in the body of your comments, that information will be posted on <https://www.regulations.gov>.

- If you want to submit a comment with confidential information that you do not wish to be made available to the public, submit the comment as a written/paper submission and in the manner detailed (see "Written/Paper Submissions" and "Instructions").

Written/Paper Submissions

Submit written/paper submissions as follows:

- *Mail/Hand delivery/Courier (for written/paper submissions):* Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852.

- For written/paper comments submitted to the Dockets Management Staff, FDA will post your comment, as well as any attachments, except for information submitted, marked and identified, as confidential, if submitted as detailed in "Instructions."

Instructions: All submissions received must include the Docket No. FDA-2004-N-0451 for "Food and Drug Administration Modernization Act of 1997: Modifications to the List of Recognized Standards, Recognition List Number: 059." Received comments will be placed in the docket and, except for those submitted as "Confidential Submissions," publicly viewable at <https://www.regulations.gov> or at the Dockets Management Staff between 9

a.m. and 4 p.m., Monday through Friday, 240-402-7500. FDA will consider any comments received in determining whether to amend the current listing of modifications to the list of recognized standards, Recognition List Number: 059.

- **Confidential Submissions**—To submit a comment with confidential information that you do not wish to be made publicly available, submit your comments only as a written/paper submission. You should submit two copies total. One copy will include the information you claim to be confidential with a heading or cover note that states "THIS DOCUMENT CONTAINS CONFIDENTIAL INFORMATION." The Agency will review this copy, including the claimed confidential information, in its consideration of comments. The second copy, which will have the claimed confidential information redacted/blacked out, will be available for public viewing and posted on <https://www.regulations.gov>. Submit both copies to the Dockets Management Staff. If you do not wish your name and contact information to be made publicly available, you can provide this information on the cover sheet and not in the body of your comments and you must identify this information as "confidential." Any information marked as "confidential" will not be disclosed except in accordance with 21 CFR 10.20 and other applicable disclosure law. For more information about FDA's posting of comments to public dockets, see 80 FR 56469, September 18, 2015, or access the information at: <https://www.govinfo.gov/content/pkg/FR-2015-09-18/pdf/2015-23389.pdf>.

Docket: For access to the docket to read background documents or the electronic and written/paper comments received, go to <https://www.regulations.gov> and insert the docket number, found in brackets in the heading of this document, into the "Search" box and follow the prompts and/or go to the Dockets Management Staff, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852, 240-402-7500.

An electronic copy of Recognition List Number: 059 is available on the internet at <https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/ucm123792.htm>. See section IV for electronic access to the searchable database for the current list of FDA-recognized consensus standards, including Recognition List Number: 059 modifications and other standards-related information. Submit written requests for a single hard copy of the document entitled "Modifications to the List of Recognized Standards, Recognition List Number: 059" to

Jianchao Zeng, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 5572, Silver Spring, MD 20993, 301-796-6580. Send one self-addressed adhesive label to assist that office in processing your request, or fax your request to 301-847-8144.

FOR FURTHER INFORMATION CONTACT:

Jianchao Zeng, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 5572, Silver Spring, MD 20993, 301-796-6580, CDRHStandardsStaff@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Section 204 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) amended section 514 of the Federal Food, Drug, and Cosmetic Act (FD&C Act) (21 U.S.C. 360d). Amended section 514 of the FD&C Act allows FDA to recognize consensus standards developed by international and national organizations for use in satisfying portions of device premarket review submissions or other requirements.

In the **Federal Register** of September 14, 2018 (83 FR 46738), FDA announced the availability of a guidance entitled "Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices." The guidance describes how FDA has implemented its standards recognition program and is available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices>. Modifications to the initial list of recognized standards, as published in the **Federal Register**, can be accessed at <https://www.fda.gov/medical-devices/standards-and-conformity-assessment-program/federal-register-documents>.

These notices describe the addition, withdrawal, and revision of certain standards recognized by FDA. The Agency maintains on its website HTML and PDF versions of the list of FDA Recognized Consensus Standards, available at <https://www.fda.gov/medical-devices/standards-and-conformity-assessment-program/federal-register-documents>. Additional information on the Agency's Standards and Conformity Assessment Program is available at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/standards-and-conformity-assessment-program>.

II. Modifications to the List of Recognized Standards, Recognition List Number: 059

FDA is announcing the addition, withdrawal, correction, and revision of certain consensus standards the Agency is recognizing for use in premarket submissions and other requirements for devices. FDA is incorporating these modifications to the list of FDA

Recognized Consensus Standards in the Agency’s searchable database. FDA is using the term “Recognition List Number: 059” to identify the current modifications.

In table 1, FDA describes the following modifications: (1) the withdrawal of standards and their replacement by others, if applicable; (2) the correction of errors made by FDA in listing previously recognized standards;

and (3) the changes to the supplementary information sheets of recognized standards that describe revisions to the applicability of the standards.

In section III, FDA lists modifications the Agency is making that involve new entries and consensus standards added as modifications to the list of recognized standards under Recognition List Number: 059.

TABLE 1—MODIFICATIONS TO THE LIST OF RECOGNIZED STANDARDS

| Old recognition No. | Replacement recognition No. | Title of standard ¹ | Change |
|---|-----------------------------|--|---|
| A. Anesthesiology | | | |
| 1–67 | 1–153 | NFPA 99:2021 Health Care Facilities Code | Withdrawn and replaced with newer version. |
| 1–78 | 1–154 | ASME PVHO–1–2019 Safety Standard for Pressure Vessels for Human Occupancy. | Withdrawn and replaced with newer version. |
| 1–132 | 1–155 | ISO 10079–2 Fourth edition 2022–03 Medical suction equipment—Part 2: Manually powered suction equipment. | Withdrawn and replaced with newer version. |
| 1–133 | 1–156 | ISO 10079–3 Fourth edition 2022–03 Medical suction equipment—Part 3: Suction equipment powered from a vacuum or positive pressure gas source. | Withdrawn and replaced with newer version. |
| 1–142 | 1–157 | ISO 10079–1 Fourth edition 2022–03 Medical suction equipment—Part 1: Electrically powered suction equipment. | Withdrawn and replaced with newer version. |
| B. Biocompatibility | | | |
| 2–93 | 2–297 | ASTM F763–22 Standard Practice for Short-Term Intramuscular Screening of Implantable Medical Device Materials. | Withdrawn and replaced with newer version. |
| 2–276 | 2–298 | ISO 10993–18 Second edition 2020–01 Amendment 1:2022–05 Biological evaluation of medical devices—Part 18: Chemical characterization of medical device materials within a risk management process [Including Amendment 1 (2022)]. | Withdrawn and replaced with newer version, including amendment. |
| 2–289 | | ISO 10993–12 Fifth edition 2021–01 Biological evaluation of medical devices—Part 12: Sample preparation and reference materials. | Transition period extended. |
| 2–296 | | ISO 10993–10 Fourth edition 2021–11 Biological evaluation of medical devices—Part 10: Tests for skin sensitization. | Transition period extended. |
| C. Cardiovascular | | | |
| No new entries at this time. | | | |
| D. Dental/Ear, Nose, and Throat | | | |
| 4–234 | 4–294 | ANSI/ADA Standard No. 139–2020 Dental Base Polymers | Withdrawn and replaced with newer version. |
| E. General I (Quality Systems/Risk Management) (QS/RM) | | | |
| 15–135 | 5–135 | ISO 20417 First edition 2021–04 Corrected version 2021–12 Medical devices—Information to be supplied by the manufacturer. | New recognition number. |
| 5–99 | 5–136 | ASTM D4332–22 Standard Practice for Conditioning Containers, Packages, or Packaging Components for Testing. | Withdrawn and replaced with newer version. |
| 5–104 | 5–137 | IEC TR 60878 Edition 4.0 2022–11 Graphical symbols for electrical equipment in medical practice. | Withdrawn and replaced with newer version. |
| 5–118 | 5–138 | AAMI TIR66:2017/(R)2020 Guidance for the creation of physiologic data and waveform databases to demonstrate reasonable assurance of the safety and effectiveness of alarm system algorithms. | New recognition number. |
| 5–119 | 5–139 | ISO 18250–3 First edition 2018–06 Medical devices—Connectors for reservoir delivery systems for healthcare applications—Part 3: Enteral application. | New recognition number. |
| F. General II (Electrical Safety/Electromagnetic Compatibility) (ES/EMC) | | | |
| 19–29 | 19–48 | IEEE ANSI/USEMCS C63.27 American National Standard for Evaluation of Wireless Coexistence. | Withdrawn and replaced with newer version. |

TABLE 1—MODIFICATIONS TO THE LIST OF RECOGNIZED STANDARDS—Continued

| Old recognition No. | Replacement recognition No. | Title of standard ¹ | Change |
|---|-----------------------------|--|---|
| G. General Hospital/General Plastic Surgery (GH/GPS) | | | |
| 6-390 | | IEC 80601-2-35 Edition 2.1 2016-04 CONSOLIDATED VERSION Medical electrical equipment-Part-2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use [Including Amendment 1 (2016)]. | Withdrawn. See 6-483. |
| 6-460 | 6-484 | ASTM F3502-22a Standard Specification for Barrier Face Coverings | Extent of recognition. Withdrawn and replaced with a newer version. |
| H. In Vitro Diagnostics (IVD) | | | |
| 7-291 | 7-313 | CLSI EP27 2nd Edition Constructing and Interpreting an Error Grid for Quantitative Measurement Procedures. | Extent of recognition. Withdrawn and replaced with newer version. |
| 7-303 | | CLSI M60 2nd Edition Performance Standards for Antifungal Susceptibility Testing of Yeast. | Withdrawn. See 7-314. |
| I. Materials | | | |
| 8-61 | 8-594 | ISO 5832-6 Third Edition 2022-03 Implants for surgery—Metallic materials—Part 6: Wrought cobalt-nickel-chromium-molybdenum alloy. | Withdrawn and replaced with newer version. |
| 8-123 | 8-595 | ISO 5832-5 Fourth Edition 2022-03 Implants for surgery—Metallic materials—Part 5: Wrought cobalt-chromium-tungsten-nickel. | Withdrawn and replaced with newer version. |
| 8-559 | 8-596 | ASTM D412-16(2021) Standard Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension. | Withdrawn and replaced with newer version. |
| J. Nanotechnology | | | |
| 18-4 | 18-21 | ISO/TS 80004-6 Second edition 2021-03 Nanotechnologies—Vocabulary—Part 6: Nano-object characterization. | Withdrawn and replaced with newer version. |
| 18-12 | 18-22 | ISO 17200 First edition 2020-09 Nanotechnology—Nanoparticles in powder form—Characteristics and measurements. | Withdrawn and replaced with newer version. |
| K. Neurology | | | |
| No new entries at this time. | | | |
| L. Obstetrics-Gynecology/Gastroenterology/Urology (OB-Gyn/G/Urology) | | | |
| 9-89 | | ISO 8638 Third edition 2010-07-01 Cardiovascular implants and extracorporeal systems—Extracorporeal blood circuit for hemodialyzers, hemodiafilters, and hemofilters. | Withdrawn. See 9-140. |
| M. Ophthalmic | | | |
| 10-37 | 10-132 | ISO 10942 Third edition 2022-01 Ophthalmic instruments—Direct ophthalmoscopes. | Extent of recognition. Withdrawn and replaced with newer version. |
| 10-91 | 10-133 | ISO 11979-10 Second edition 2018-03 Ophthalmic implants—Intraocular lenses—Part 10: Clinical investigations of intraocular lenses for correction of ametropia in phakic eyes. | Withdrawn and replaced with newer version. |
| N. Orthopedic | | | |
| 11-264 | 11-394 | ASTM F1820-22 Standard Test Method for Determining the Forces for Disassembly of Modular Acetabular Devices. | Withdrawn and replaced with newer version. |
| 11-306 | 11-395 | ASTM F1814-22 Standard Guide for Evaluating Modular Hip and Knee Joint Components. | Withdrawn and replaced with newer version. |
| 11-320 | 11-396 | ISO 7206-13 First edition 2016-07-01 [Including AMD1:2022] Implants for surgery—Partial and total hip joint prostheses—Part 13: Determination of resistance to torque of head fixation of stemmed femoral components [Including Amendment 1 (2022)]. | Withdrawn and replaced with newer version including amendment. |
| O. Physical Medicine | | | |
| 16-191 | | ISO 7176-16 Second edition 2012-12-01 Wheelchairs—Part 16: Resistance to ignition of postural support devices. | Withdrawn. See 16-233. |
| P. Radiology | | | |
| 12-113 | 12-346 | ISO 12005 Third edition 2022-05 Lasers and laser-related equipment—Test methods for laser beam parameters—Polarization. | Withdrawn and replaced with newer version. |

TABLE 1—MODIFICATIONS TO THE LIST OF RECOGNIZED STANDARDS—Continued

| Old recognition No. | Replacement recognition No. | Title of standard ¹ | Change |
|--------------------------------|-----------------------------|--|--|
| 12-295 | 12-347 | IEC 60601-2-33 Edition 4.0 2022-08 Medical electrical equipment—Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis. | Extent of recognition. Withdrawn and replaced with newer version. |
| 12-317 | 12-348 | IEC 60601-2-54 Edition 2.0 2022-09 Medical electrical equipment—Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy. | Extent of recognition. Withdrawn and replaced with newer version. |
| 12-342 | 12-349 | NEMA Digital Imaging and Communications in Medicine (DICOM) Set PS3.1-3.20 2022d. | Withdrawn and replaced with newer version. |
| Q. Software/Informatics | | | |
| 13-109 | 13-121 | ANSI/AAMI/UL 2800-1:2022 Standard for Medical Device Interoperability | Withdrawn and replaced with newer version. See 13-125, 13-126, 13-127. |
| R. Sterility | | | |
| 14-409 | 14-580 | ISO 11137-2 Third edition 2013-06 [Including AMD1:2022] Sterilization of health care products—Radiation—Part 2: Establishing the sterilization dose [Including Amendment 1 (2022)]. | Withdrawn and replaced with newer version. |
| 14-527 | 14-581 | ASTM F2638-22 Standard Test Method for Using Aerosol Filtration for Measuring the Performance of Porous Packaging Materials as a Surrogate Microbial Barrier. | Withdrawn and replaced with newer version. |
| S. Tissue Engineering | | | |
| No new entries at this time. | | | |

¹ All standard titles in this table conform to the style requirements of the respective organizations.

III. Listing of New Entries

In table 2, FDA provides the listing of new entries and consensus standards added as modifications to the list of recognized standards under Recognition List Number: 059. These entries are of standards not previously recognized by FDA.

TABLE 2—NEW ENTRIES TO THE LIST OF RECOGNIZED STANDARDS

| Recognition No. | Title of standard ¹ | Reference No. and date |
|------------------------------|---|---------------------------------------|
| A. Anesthesiology | | |
| 1-158 | Medical suction equipment—Part 4: General requirements | ISO 10079-4 First edition 2021-08. |
| 1-159 | Respiratory equipment—Particular requirements for basic safety and essential performance of infant cardiorespiratory monitors. | ISO 18778 Second edition 2022-06. |
| 1-160 | Medical electrical equipment—Part 2-84: Particular requirements for the basic safety and essential performance of ventilators for the emergency medical services environment. | ISO 80601-2-84 First edition 2020-07. |
| B. Biocompatibility | | |
| No new entries at this time. | | |
| C. Cardiovascular | | |
| 3-183 | Cardiovascular implants and extracorporeal systems—Blood/tissue contact surface modifications for extracorporeal perfusion systems. | ISO 11658 First edition 2012-05-15. |
| D. Dental/ENT | | |
| 4-295 | Evaluation of biocompatibility of medical devices used in dentistry | ANSI/ADA Standard No. 41-2020. |
| 4-296 | Dentistry—Intra-oral mirrors | ISO 9873 Fourth edition 2019-03. |
| 4-297 | Dentistry—Manual toothbrushes—General requirements and test methods | ISO 20126 Third edition 2022-03. |
| E. General I (QS/RM) | | |
| 5-140 | Standard for verification and validation in computational solid mechanics | ASME V&V 10-2019. |
| 5-141 | Standard for verification and validation in computational fluid dynamics and heat transfer. | ASME V&V 20-2009 (R2021). |

TABLE 2—NEW ENTRIES TO THE LIST OF RECOGNIZED STANDARDS—Continued

| Recognition No. | Title of standard ¹ | Reference No. and date |
|--------------------------------|--|--|
| F. General II (ES/EMC) | | |
| No new entries at this time. | | |
| G. GH/GPS | | |
| 6–483 | Medical electrical equipment—Part 2–35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads and mattresses and intended for heating in medical use. | IEC 60601–2–35 Edition 2.0 2020–09. |
| 6–485 | Sterile hypodermic syringes for single use—Part 4: Syringes with re-use prevention feature. | ISO 7886–4 Second Edition 2018–11. |
| H. IVD | | |
| 7–314 | Performance Standards for Antifungal Susceptibility Testing of Yeasts | CLSI M27M44S, 3rd Edition. |
| I. Materials | | |
| No new entries at this time. | | |
| J. Nanotechnology | | |
| No new entries at this time. | | |
| K. Neurology | | |
| No new entries at this time. | | |
| L. OB-Gyn/G/Urology | | |
| 9–140 | Extracorporeal systems for blood purification—Part 2: Extracorporeal blood circuit for hemodialyzers, hemodiafilters and hemofilters. | ISO 8637–2 First Edition 2018–07. |
| 9–141 | Extracorporeal systems for blood purification—Part 3: Plasmafilters | ISO 8637–3 First Edition 2018–07. |
| 9–142 | Standard test method for static and kinetic coefficients of friction of plastic film and sheeting. | ASTM D1894–14. |
| 9–143 | Sterile urethral catheters for single use | ISO 20696 First edition 2018–06 Corrected 2019–12. |
| 9–144 | Sterile drainage catheters and accessory devices for single use | ISO 20697 First edition 2018–06 Corrected 2019–09. |
| M. Ophthalmic | | |
| No new entries at this time. | | |
| N. Orthopedic | | |
| 11–397 | Standard test method for fatigue testing of total knee femoral components under closing conditions. | ASTM F3210–22e1. |
| 11–398 | Standard test methods for sacroiliac joint fusion devices | ASTM F3574–22. |
| O. Physical Medicine | | |
| 16–233 | Wheelchair seating—Part 10: Resistance to ignition of postural support devices—Requirements and test method. | ISO 16840–10 Second edition 2021–06 Corrected version 2022–01. |
| P. Radiology | | |
| No new entries at this time. | | |
| Q. Software/Informatics | | |
| 13–122 | Health software and health IT systems safety, effectiveness and security—Part 5–1: Security—Activities in the product life cycle. | IEC 81001–5–1 Edition 1.0 2021–12. |
| 13–123 | Manufacturer disclosure statement for medical device security | ANSI/NEMA HN 1–2019. |
| 13–124 | Guidance on the application of ISO 14971 to artificial intelligence and machine learning. | AAMI CR34971:2022. |
| 13–125 | Standard for risk concerns for interoperable medical products | ANSI/AAMI/UL 2800–1–1:2022. |
| 13–126 | Standard for interoperable item development life cycle | ANSI/AAMI/UL 2800–1–2:2022. |
| 13–127 | Standard for Interoperable item integration life cycle | ANSI/AAMI/UL 2800–1–3:2022. |
| 13–128 | IEEE/UL Standard for wireless diabetes device security: Information security requirements for connected diabetes solutions. | IEEE Std 2621.2–2022/UL 2621–2:2022. |

TABLE 2—NEW ENTRIES TO THE LIST OF RECOGNIZED STANDARDS—Continued

| Recognition No. | Title of standard ¹ | Reference No. and date |
|------------------------------|---|---------------------------------------|
| R. Sterility | | |
| 14–582 | Sterilization of health care products—Radiation—Part 4: Guidance on process control | ISO/TS 11137–4 First edition 2020–06. |
| 14–583 | Cleaning validation of health care products—Requirements for development and validation of a cleaning process for medical devices.. | ANSI/AAMI ST98:2022. |
| S. Tissue Engineering | | |
| No new entries at this time. | | |

¹ All standard titles in this table conform to the style requirements of the respective organizations.

IV. List of Recognized Standards

FDA maintains the current list of FDA Recognized Consensus Standards in a searchable database that may be accessed at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>. Such standards are those that FDA has recognized by notice published in the **Federal Register** or that FDA has decided to recognize but for which recognition is pending (because a periodic notice has not yet appeared in the **Federal Register**). FDA will announce additional modifications and revisions to the list of recognized consensus standards, as needed, in the **Federal Register** once a year, or more often if necessary.

V. Recommendation of Standards for Recognition by FDA

Any person may recommend consensus standards as candidates for recognition under section 514 of the FD&C Act by submitting such recommendations, with reasons for the recommendation, to CDRHStandardsStaff@fda.hhs.gov. To be considered, such recommendations should contain, at a minimum, the information available at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/standards-and-conformity-assessment-program#process>.

Dated: July 28, 2023.

Lauren K. Roth,

Associate Commissioner for Policy.

[FR Doc. 2023–16418 Filed 8–1–23; 8:45 am]

BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA–2023–N–2780]

Agency Information Collection Activities; Proposed Collection; Comment Request; Premarket Notification for a New Dietary Ingredient

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA or Agency) is announcing an opportunity for public comment on the proposed collection of certain information by the Agency. Under the Paperwork Reduction Act of 1995 (PRA), Federal Agencies are required to publish notice in the **Federal Register** concerning each proposed collection of information and to allow 60 days for public comment in response to the notice. This notice solicits comments on the procedure by which a manufacturer or distributor of a new dietary ingredient or of a dietary supplement containing a new dietary ingredient is to submit to FDA information upon which it has based its conclusion that a dietary supplement containing the new dietary ingredient will reasonably be expected to be safe. **DATES:** Either electronic or written comments on the collection of information must be submitted by October 2, 2023.

ADDRESSES: You may submit comments as follows. Please note that late, untimely filed comments will not be considered. The <https://www.regulations.gov> electronic filing system will accept comments until 11:59 p.m. Eastern Time at the end of October 2, 2023. Comments received by mail/hand delivery/courier (for written/paper submissions) will be considered timely if they are received on or before that date.

Electronic Submissions

Submit electronic comments in the following way:

- **Federal eRulemaking Portal:** <https://www.regulations.gov>. Follow the instructions for submitting comments. Comments submitted electronically, including attachments, to <https://www.regulations.gov> will be posted to the docket unchanged. Because your comment will be made public, you are solely responsible for ensuring that your comment does not include any confidential information that you or a third party may not wish to be posted, such as medical information, your or anyone else's Social Security number, or confidential business information, such as a manufacturing process. Please note that if you include your name, contact information, or other information that identifies you in the body of your comments, that information will be posted on <https://www.regulations.gov>.

- If you want to submit a comment with confidential information that you do not wish to be made available to the public, submit the comment as a written/paper submission and in the manner detailed (see “Written/Paper Submissions” and “Instructions”).

Written/Paper Submissions

Submit written/paper submissions as follows:

- **Mail/Hand Delivery/Courier (for written/paper submissions):** Dockets Management Staff (HFA–305), Food and Drug Administration, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852.

- For written/paper comments submitted to the Dockets Management Staff, FDA will post your comment, as well as any attachments, except for information submitted, marked and identified, as confidential, if submitted as detailed in “Instructions.”

Instructions: All submissions received must include the Docket No. FDA–2023–N–2780 for “Agency Information Collection Activities; Proposed Collection; Comment Request; Premarket Notification for a New