

DEPARTMENT OF HEALTH AND HUMAN SERVICES**Food and Drug Administration****21 CFR Part 868****[Docket No. FDA-2021-N-0917]****Medical Devices; Anesthesiology Devices; Classification of the Retrograde Intubation Device****AGENCY:** Food and Drug Administration, HHS.**ACTION:** Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA or we) is classifying the retrograde intubation device into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for the retrograde intubation device's classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices.

DATES: This order is effective December 28, 2021. The classification was applicable on December 12, 2018.

FOR FURTHER INFORMATION CONTACT: Todd Courtney, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 1216, Silver Spring, MD 20993-0002, 301-796-4634, Todd.Courtney@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:**I. Background**

Upon request, FDA has classified the retrograde intubation device as class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness. In addition, we believe this action will enhance patients' access to beneficial innovation, by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified as, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (see 21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in

commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (see 21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k) and part 807 (21 CFR part 807)).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act. Section 207 of the Food and Drug Administration Modernization Act of 1997 established the first procedure for De Novo classification (Pub. L. 105-115). Section 607 of the Food and Drug Administration Safety and Innovation Act modified the De Novo application process by adding a second procedure (Pub. L. 112-144). A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a 510(k) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval

application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less-burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On September 25, 2017, FDA received Cook Incorporated's request for De Novo classification of the Retrograde Intubation Set. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see 21 U.S.C. 360c(a)(1)(B)). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on December 12, 2018, FDA issued an order to the requester classifying the device into class II. In this final order, FDA is codifying the classification of the device by adding 21 CFR 868.5095.¹ We have named the generic type of device retrograde intubation device, and it is identified as a prescription device used to perform retrograde intubation via the cricothyroid membrane. The device may contain or be labeled for use with guidewires and intubating catheters, in addition to needles (21 CFR 868.5090), syringe (21 CFR 880.5860), and hemostats (21 CFR 878.4800).

FDA has identified the following risks to health associated specifically with this type of device and the measures

¹ FDA notes that the "ACTION" caption for this final order is styled as "Final amendment; final order," rather than "Final order." Beginning in December 2019, this editorial change was made to indicate that the document "amends" the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register's (OFR) interpretations of the Federal Register Act (44 U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

required to mitigate these risks in table 1.

TABLE 1—RETROGRADE INTUBATION DEVICE RISKS AND MITIGATION MEASURES

Identified risks	Mitigation measures
Failure to intubate and ventilate (continued hypoxia)	Non-clinical performance testing, and Labeling
Tissue damage/trauma resulting in, for example:	Non-clinical performance testing, and Labeling
• Bleeding, hematoma	
• Subcutaneous emphysema	
• Pneumomediastinum or pneumothorax	
• Damage to trachea, esophagus, and vocal cords	
Infection	Sterilization validation, and Shelf-life testing
Adverse tissue reaction	Biocompatibility evaluation

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this order. This device is subject to premarket notification requirements under section 510(k) of the FD&C Act.

At the time of classification, retrograde intubation devices are for prescription use only. Prescription devices are exempt from the requirement for adequate directions for use for the layperson under section 502(f)(1) of the FD&C Act (21 U.S.C. 352(f)(1)) and 21 CFR 801.5, if the conditions of 21 CFR 801.109 are met.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

While this final order contains no collection of information, it does refer to previously approved FDA collections of information. Therefore, clearance by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (PRA) (44 U.S.C. 3501–3521) is not required for this final order. The previously approved collections of information are subject to review by OMB under the PRA. The collections of information in the guidance document “De Novo Classification Process (Evaluation of Automatic Class III Designation)” have been approved under OMB control number 0910–0844; the collections of information in 21 CFR part 814, subparts A through E,

regarding premarket approval, have been approved under OMB control number 0910–0231; the collections of information in part 807, subpart E, regarding premarket notification submissions, have been approved under OMB control number 0910–0120; the collections of information in 21 CFR part 820, regarding quality system regulation, have been approved under OMB control number 0910–0073; and the collections of information in 21 CFR part 801, regarding labeling, have been approved under OMB control number 0910–0485.

List of Subjects in 21 CFR Part 868

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 868 is amended as follows:

PART 868—ANESTHESIOLOGY DEVICES

■ 1. The authority citation for part 868 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

■ 2. Add § 868.5095 to subpart F to read as follows:

§ 868.5095 Retrograde intubation device.

(a) *Identification.* A retrograde intubation device is a prescription device used to perform retrograde intubation via the cricothyroid membrane. The device may contain or be labeled for use with guidewires and intubating catheters, in addition to needles (§ 868.5090), syringe (§ 880.5860 of this chapter), and hemostats (§ 878.4800 of this chapter).

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use, including the following:

- (i) Wire guide tensile, flex, fracture, and corrosion testing;
- (ii) Catheter tensile strength testing at likely points of failure;
- (iii) Catheter kink radius testing;
- (iv) Compatibility of device components that interact, including compatibility in connection, disconnection, and ability to transfer fluids;
- (v) Dimensional validation;
- (vi) Accuracy testing of markings; and
- (vii) Validation of the maximum airway pressure.

(2) Performance data must support the shelf life of the device by demonstrating continued sterility, package integrity, and device functionality over the identified shelf life.

(3) The device must be demonstrated to be biocompatible.

(4) Labeling must include:

- (i) Instructions for use; and
- (ii) Package labels that clearly identify the minimum compatible size of endotracheal tube.

Dated: December 16, 2021.

Lauren K. Roth,

Associate Commissioner for Policy.

[FR Doc. 2021–28166 Filed 12–27–21; 8:45 am]

BILLING CODE 4164–01–P

GENERAL SERVICES ADMINISTRATION

41 CFR Parts 300–3, 302–2, 302–3, 302–12, 302–15, and 302–17

[FTR Case 2020–302–1; Docket No. 2020–0019, Sequence 1]

RIN 3090–AK31

Federal Travel Regulation; Taxes on Relocation Expenses, Withholding Tax Allowance (WTA) and Relocation Income Tax Allowance (RITA) Eligibility

AGENCY: Office of Government-wide Policy (OGP), General Services Administration (GSA).

ACTION: Final rule.